



CERTIFICATION

AOAC Research Institute *Performance Tested Methods*SM

Certificate No.
021402

The AOAC Research Institute hereby certifies the method known as:

HistaSure™ ELISA^{Fast Track}

manufactured by

**Labor Diagnostika Nord GmbH & Co
KG Am Eichenhain 1
48531 Nordhorn
Germany**

This method has been evaluated in the AOAC Research Institute *Performance Tested Methods*SM Program and found to perform as stated in the applicability of the method. This certificate indicates an AOAC Research Institute Certification Mark License Agreement has been executed which authorizes the manufacturer to display the AOAC Research Institute *Performance Tested Methods*SM certification mark on the above-mentioned method for the period below. Renewal may be granted by the Expiration Date under the rules stated in the licensing agreement.

A handwritten signature in black ink that reads "Scott Coates".

Scott Coates, Senior Director
Signature for AOAC Research Institute

Issue Date	October 21, 2023
Expiration Date	December 31, 2024

AUTHORS

Georg Manz and Esther Bootlink

SUBMITTING COMPANY

LDN Labor Diagnostika Nord GmbH & Co. KG
 Am Eichenhain 1
 48531 Nordhorn
 Germany

KIT NAME(S)HistaSure™ ELISA ^{Fast Track}**CATALOG NUMBERS**

FC E-3600, FC E-3900

INDEPENDENT LABORATORY

Eurofins Central Analytical Laboratories
 2315 N. Causeway Blvd.
 Metairie, LA 70001
 USA

APPLICABILITY OF METHOD

Target analyte– histamine.

Matrixes – (10 g samples) - fresh/frozen yellowfin tuna, canned tuna-chunk light in water, fresh/frozen mahi mahi, canned sardines in oil, and fish meal

Performance claims

Selectivity: Cross reactivity with components of similar behavior < 2%

Accuracy: %RSD < 15% (above LOQ)

Precision: recovery within 80 – 120 %

Detection limit: LOQ < 3 ppm (Corresponding to Control 2 of the ELISA)

Stability: 2-year shelf life at 2 – 8°C. Acceptable OD range of control 0: 1.250 – 2.750.

Method comparison LDN HistaSure ELISAFast Track vs AOAC 977.13.

Correlation Coefficient $r > 0.9$. Ratio AOAC 977.13 / HistaSure ELISA^{Fast Track} 0.8 – 1.2**REFERENCE METHOD**

W. Horwitz, G.W. Latimer, 2005. AOAC Official Method 977.13. Official Methods of analysis of AOAC International 18th Edition 35.1.32 (13)

ORIGINAL CERTIFICATION DATE

February 03, 2014

CERTIFICATION RENEWAL RECORD

Renewed annually through December 2024.

METHOD MODIFICATION RECORD

1. November 2017 Level 1
2. November 2019 Level 1
3. September 2020 Level 1
4. December 2022 Level 1
5. October 2023 Level 1

SUMMARY OF MODIFICATION

1. Editorial changes to inserts.
2. Editorial changes to inserts.
3. Editorial changes to labels.
4. Editorial changes to inserts.
5. Editorial changes to inserts.

Under this AOAC *Performance Tested Methods*SM License Number, 021402 this method is distributed by:
 NONE

Under this AOAC *Performance Tested Methods*SM License Number, 021402 this method is distributed as:
 NONE

PRINCIPLE OF THE METHOD (1)

Histamine is extracted from the fish using a simple water extraction. After filtration or centrifugation, the clear supernatant is used for the subsequent step, the derivatization of histamine.

By use of the acylation reagent, histamine from the calibrators and histamine from the food extracts is quantitatively derivatized into N-acylhistamine. The competitive Histamine ELISA kit uses the microtiter plate format. Histamine is bound to the solid phase of the microtiter plate. Acylated histamine and solid phase bound histamine compete for a fixed number of antiserum binding sites. When the system is in equilibrium, free antigen and free antigen-antiserum-peroxidase complexes are removed by washing. The substrate TMB/peroxidase reaction is monitored at 450 nm. The amount of antibody bound to the solid phase histamine is inversely proportional to the histamine concentration of the sample.

DISCUSSION OF THE VALIDATION STUDY (1)

In the AOAC PTM validation study the performance of the LDN HistaSure™ ELISA^{Fast Track} was at least equivalent to the AOAC 977.13 reference method. The LDN HistaSure ELISA^{Fast Track} showed good recovery rates for all tested matrixes (fresh/frozen tuna, canned tuna, fresh/frozen mahi mahi, canned sardines and fish meal) and repeatability precision was excellent at the 50 ppm defect action level. No solvents are required for extraction. The method is robust, fast, efficient and suitable for high throughput screening. Based on the results of this study it could be recommended that the LDN HistaSure ELISA^{Fast Track} receives PTM certification.

Table 9: Method comparison results overview AOAC 977.13 vs. LDN HistaSure™ ELISA^{Fast Track} (1)

	Fortification level (ppm)	AOAC 977.13				LDN HistaSure™ ELISA ^{Fast Track}			
		Mean concentration (n=5) (ppm)	SD	%RSD	Average % recovery (n=5)	Mean concentration (n=7) (ppm)	SD	%RSD	Average % recovery (n=7)
Fresh/frozen Tuna	0	0.51	0.63	123.5	/	0.27 (< LOQ)	0.11	38.7	/
	5	4.92	0.81	16.4	88.1	5.25	0.40	7.52	99.6
	20	17.6	2.81	16.0	85.4	19.2	2.08	10.8	94.6
	50	43.9	7.37	16.8	86.8	42.9	1.77	4.13	85.2
	100	87.9	15.1	17.2	87.4	92.6	6.99	7.55	92.4
	250	261.8	27.2	10.4	104.5	218.6	17.3	7.90	87.3
	Mean recovery				90.4				91.8
Canned tuna	0	0.22	0.22	99.7	/	0.50 (< LOQ)	0.17	33.7	/
	5	4.84	0.16	3.21	92.4	5.49	0.76	13.9	99.8
	20	18.3	2.35	12.8	90.6	20.2	2.01	9.92	98.6
	50	52.3	2.89	5.52	104.2	47.8	4.10	8.57	94.7
	100	99.5	7.12	7.15	99.3	99.8	8.10	8.12	98.7
	250	238.9	33.8	14.2	95.5	267.3	8.46	3.17	106.5
	Mean recovery				96.4				99.8
Frozen mahi mahi	0	1.52	0.96	63.3	/	1.22 (< LOQ)	0.21	17.3	/
	5	5.44	0.77	14.3	78.4	6.38	0.61	9.56	103.1
	20	17.9	1.69	9.45	82.0	18.9	1.28	6.78	88.2
	50	45.8	7.68	16.8	88.5	43.5	2.29	5.26	84.6
	100	93.9	6.30	6.71	92.4	82.7	2.46	2.97	81.5
	250	242.6	19.9	8.22	96.4	199.0	12.6	6.35	79.1
	Mean recovery				87.5				87.3
Canned sardines	0	0.21	0.32	155.1	/	0.67 (< LOQ)	0.19	28.9	/
	5	4.24	0.28	6.57	80.6	5.27	0.28	5.36	92.1
	20	20.7	1.10	5.32	102.3	18.2	1.24	6.78	87.8
	50	41.6	4.89	11.8	82.9	41.5	0.75	1.80	81.6
	100	86.1	10.8	12.5	85.9	77.5	3.94	5.08	76.8
	250	230.3	32.2	14.0	92.0	249.9	23.0	9.22	99.7
	Mean recovery				89.0				87.6
Fish meal	0	6.31	0.73	11.6	/	9.4	0.54	5.71	/
	5	5.30	0.30	5.73	-20.1	14.0	1.11	7.92	92.3
	20	15.5	1.42	9.12	46.2	25.5	1.02	3.98	80.5
	50	38.5	4.11	10.7	64.5	48.9	1.02	2.09	79.0
	100	95.1	2.78	2.92	88.8	93.8	3.73	3.97	84.4
	250	212.9	19.8	9.29	82.6	244.2	15.8	6.47	93.9
	Mean recovery				52.4				86.0
Naturally contaminated	Sample 1	61	3.06	5.04	/	58	5.2	8.95	/
	Sample 2	203	13.7	6.74	/	133	10.7	7.99	/
	Sample 3	7	1.00	14.3	/	8.6	0.92	10.7	/
	Sample 4	54	3.21	5.99	/	48	2.8	5.83	/
	Sample 5	57	1.53	2.66	/	69	5.3	7.75	/

REFERENCES CITED

1. Manz, Georg, and Booltink, Esther., Evaluation of the HistaSure™ ELISA ^{Fast Track} for Determination of Histamine in Fish Samples, *AOAC Performance Tested MethodsSM* certification number 021402.
2. AOAC Research Institute Validation Outline for HistaSure™ ELISA ^{Fast Track} , Approved –February 2014.
3. P.L. Rogers, W.F. Staruszkiewicz, 2000. Histamine Test Kit Comparison. *Journal of Aquatic Food Product Technology* 9 (2), 5-17
4. J.M. Hungerford, 2010. Scombroid poisoning: a review. *Toxicol* 15, 231-243
5. S. Köse, N. Kaklkkaya, S. Koral, B. Tufan, K.C. Buruk, F. Aydin, 2011. Commercial test kits and the determination of histamine in traditional (ethnic) fish products-evaluation against EU accepted HPLC method. *Food Chemistry*, 125 (4) 1490-1497
6. A. Lupo, M. Mozola, 2011. Validation study of a rapid ELISA for the detection of histamine in tuna. *Journal of AOAC International* 94 (3), 886-899
7. D.G. Altman. J.M. Bland, 1983. Measurement in medicine: the analysis of method comparison studies. *The Statistician* 32, 307-317
8. R. Taylor. 1990. Interpretation of the correlation coefficient: A basic review. *Journal of diagnostic medical sonography* 6(1), 35-39
9. National Association of Testing Authorities, Australia, 2012. Technical Note 17: Guidelines for the validation and verification of quantitative and qualitative test methods.
10. US department of health and human services food and drug administration, 2001.Guidance for Industry, bioanalytical method validation.
11. M. Thompson, S.L.R. Ellison, R. Wood, 2002. Harmonized guidelines for single laboratory validation of methods analysis. *Pure and applied chemistry* 74, 835-855
12. NCCLS, Protocols for Determination of Limits of Detection and Limits of Quantification; Approved Guideline. EP 17-A Vol. 24 No. 10
13. W. Horwitz, G.W. Latimer, 2005. AOAC Official Method 977.13. Official Methods of analysis of AOAC International 18th Edition 35.1.32