













15. Morganti, A; Lonati, C; Turolo, L (2010) Brief Cryoactivation Markedly Affects Plasma Renin Activity and Direct Renin Measurement in Low Renin Samples: PP.24.473, Journal of Hypertension: 28 p e388 doi: 10.1097/01.hjh.0000379399.50039.3c.
16. Sealey JE. (1991) Plasma renin activity and plasma prorenin assays. Clin Chem. 37(10 Pt 2):1811-9. PMID: 1914195.
17. Campbell, D. J., Nussberger, J., Stowasser, M., Danser, A. H., Morganti, A., Frandsen, E., & Ménard, J. (2009). Activity assays and immunoassays for plasma Renin and prorenin: information provided and precautions necessary for accurate measurement. Clinical chemistry, 55(5), 867–877.
18. Brossaud J, Corcuff JB. Pre-Analytical and Analytical Considerations for the Determination of Plasma Renin Activity. Clin Chim Acta. 2009; 410(1–2):90–2.

18. CHANGE HISTORY

Previous Version:	6.0-r	New Version:	7.0-r
Changes:	Design change of product and new IFU format; all information in IFU was revised.		
Previous Version:	7.0-r	New Version:	8.0-r
	<p>7.2 Specimen Pre-Treatment & Storage Angiotensin-I Generation Procedure</p> <p>8. Added: If the generated samples will be tested immediately, ... The samples are now ready for testing.</p> <p>9. Added: If the generated samples will be tested at a later time, ... Added information on how to bring the frozen generated samples to room temperature prior to testing.</p> <p>9. Reagents Provided</p> <p>9. Stopping Solution</p> <p>The hazard pictogram GHS05 has been replaced by the hazard pictogram GHS07.</p>		

Symbols:

	Storage temperature		Manufacturer		Contains sufficient for <n> tests
	Use-by date		Batch code		
	Consult instructions for use		Content		
	Caution		Catalogue number		Distributor
	Date of manufacture				For research use only!